

StaebeL Total Flow Hybrid Air Mattress User Manual

Product Code: 3CM-CMBR



Thank you for choosing Staebel Total Hybrid Air Mattress

We aim to ensure you are satisfied with your mattress in terms of both performance and durability. We therefore kindly ask you to spend a few minutes reading through these user instructions to ensure you can take the best possible care of your mattress.

If you have any questions that are not answered in these user instructions, please contact us at Staebel. You will find our contact details on the back page of these user instructions.

OPERATING PRINCIPLES

- The Total Flow-Hybrid can be used just as a high specification foam pressure redistribution mattress, and also functions as a balanced-air self-adjusting system with foam-filled air cells, and the option to add a choice of pumps to provide alternating pressure relief, and a built in Total Evacuation System of handles and straps.
- Use as static or dynamic system, no requirement to remove patient from bed to change from foam system to alternating air system. Total Flow-Hybrid also boosts a unique adaptable heel zone with gel-foam pads, individually removable to reduce to zero pressure on the heel as required for unrivaled pressure redistribution and improved circulation.
- Alternating pressure at a sufficient rhythm avoids prolonged vascular compression which is likely to cause tissue hypoxia.
- The patient is an intended operator. The patient can use the device and its accessories according to this manual.

GUIDELINES FOR USE

- The traditionally accepted method of pressure sore control is to turn the client at regular intervals and to observe the areas of risk regularly.
- Alternating pressure mattress is intended to complement this treatment, Observation of the client and turning when necessary should still be considered as part of the treatment.
- Always consult a physician or health professional before using this device.

CONTROL PANEL FEATURES

- Pressure adjustable knob easy operation for setup of patient weight (max patient weight setting 250kgs).
- Auto pressure control through air sensor and microcomputer control.
- 10mins, 15mins, 20mins, 25mins cycle time options.
- Static, alternate, max firm and lock modes.
- Low pressure, power failure and alternate failure warning lights and alarm.



INSTALLATION/USE

1. Unpack the mattress and place on bed with the feet logo at the foot end.
2. Hang the compressor (pump) on the bed using the hooks.
3. Connect 2-pipe connector to end of mattress
4. Tubing (supplied in pump box)
5. Connect the mattress feed pipes to the compressor (pump) connector by matching the appropriate outlets.
6. Connect the electrical power supply cable to the compressor (pump).
7. Guide the electrical power supply cable from the compressor (pump) along the bed to the plug socket and plug in.
8. Press switch to turn on the compressor (pump).
9. The pump will enter static mode initially until pressure reaches 15mmHg (about 30 minutes), then it will change to alternate mode and keep inflating the mattress to the configured pressure. Once configured pressure is reached, the pump will start functioning normally.
10. Select the pressure corresponding to the patient's morphology.
11. For nursing and caring convenience, press the max firm button to automatically inflate the mattresses to the maximum level. The pressure will return to alternate mode after 20 minutes.
12. Max firm inflates whole mattress to max level. Static is different, static ceases alternate function but allows pressure to still be selected.

LOCK BUTTON

1. **Auto:** Control panel will automatically lock in 5 minutes without operation.
2. **Manual:** press lock button for 2 seconds to lock the panel, press again for 2 seconds to unlock the panel.

MEMORY FUNCTION AFTER POWER FAILURE

If the power supply is stopped and the switch is ON position, when power resumes, the pump will resume to the configured mode before power failure. However if the pump is in Static or Max Firm mode, the pump will be reverted to Alternate Mode after re-switching on the pump.

EMERGENCY CPR DEFLATION

In the event of an emergency, remove the mattress pipe connector block from the pump and turn off the compressor (pump). The air will discharge from the mattress with the patient's own weight. To resume normal operation follow the above steps.

SAFETY MEASURES

- DO TAKE CARE when operating this electrical equipment.
- DO NOT leave connected to electrical supply when not in use.
- DO NOT expose the equipment to excessive humidity.
- DO NOT use corrosive cleaning products such as industrial degreasers or acetic solvents.
- DO NOT attempt to relocate this equipment whilst connected to the electrical supply.
- DO NOT position the equipment to make it difficult to operate the power plug used to isolate the equipment circuits electrically from the mains supply.
- **WARNING:** To avoid the risk of electric shock, this equipment must only be connected to a mains supply with protective earth.

CLEANING/DISINFECTION

- Air mattress system must be cleaned thoroughly between patients to avoid cross contamination.
- DO TAKE CARE to disconnect from the electrical supply before attempting to clean.
- DO TAKE CARE to use a cloth saturated with a mild detergent solution.
- DO TAKE CARE to follow manufacturers solution concentrations
- DO TAKE CARE not to exceed manufacturer’s recommendations before wiping off cleaning solution.
- Clean using a single use wipe or damp cloth and a solution of mild surface detergent/ disinfectant with CE mark and neutral water. Rinse thoroughly with clean water. If the cover is heavily soiled a 0.1% (1,000ppm) chlorine solution and cold water can be used. Where required, covers can be laundered in the washing machine as follows; main wash 65°C for not less than 10 minutes or 95°C for not less than 3 minutes.
- Drying Covers can be tumble dried on low heat setting. Drying temperature must not exceed 60°C.

SERVICE & WARRANTY

This device is not self-serviceable.

Maintenance must ONLY be performed by an authorised Technician. Failure to use genuine Total parts will void the warranty of this product.

We warrant the product to be free from manufacturing defects for the period of **12 months** (mattress) and **24 months** (pump) from date of purchase. We will make available upon request; component part lists, descriptions, or other information that will assist service personnel to repair those parts of ME equipment that are designated by the manufacturer as repairable by service personnel.

ENVIRONMENTAL PROTECTION

The device and accessories are to be disposed of according to local regulations after their service lives. Alternatively, they can be returned to the dealer or the manufacturer for recycling or proper disposal. Max firm inflates whole mattress to max level. Static is different, static ceases alternate function but allows pressure to still be selected.

SPECIFICATIONS

Mattress Specifications		
Size	Single	King Single
Mattress Dimensions	1970 x 900 x 150 mm	1970 x 1070 x 150 mm
Materials	Nylon TPU cell, nylon PU multi-way stretch cover & CPR	
Max. User Weight	30 - 250 kg	
Pillow function		
Warranty	12 months unless Otherwise specified	
Standards	ISO 9001:2015 ISO 13485:2016	

Pump Specifications	
Cycle Time	10/15/20/25 mins
Air Output	SL/min
Electrical Power Supply	220-240V -SCMOHz
Warranty	24 months unless Otherwise specified